



INDIAN PHARMACOPOEIA COMMISSION

MINISTRY OF HEALTH & FAMILY WELFARE, GOVERNMENT OF INDIA

SECTOR-23, RAJ NAGAR, GHAZIABAD- 201 002.

Tel No: 0120- 2783392, 2783400, 2783401; Fax: 2783311

Mail: lab.ipc@gov.in, Web: www.ipc.gov.in

Dated: 09.10.2020

Corrigendum- I

Sub: Corrigendum to the Tender Document for Procurement of "KARL FISCHER TITRATOR" at IPC, Ghaziabad against Tender No. IPC/5519/2020-21, dated 18.08.2020- reg.

The following changes are being incorporated in above referred Tender under sections:-

1. SECTION-II (TECHNICAL SPECIFICATIONS) and SECTION-VI (D) TECHNICAL COMPLIANCE SHEET:

FOR:

Sr. No	Parameter/ Particular	Tendered Specifications
3	mV range	+/- 3200mV
6	Polarized Measurement Range	0 to +/- 3200mV or better
10	Measurement range	10 ppm- 100%

READ AS:

Sr. No	Parameter/ Particular	Amended Specifications
3	mV range	+/- 1200mV
6	Polarized Measurement Range	+/- 1200mV
10	Measurement range	50 ppm- 100%
Added Specification Sr. No. 20	21 CFR Part 11 compliance for Equipment	Must comply

2. SECTION-I (NOTICE INVITING TENDER) Important dates have been changed as:-

LAST DATE FOR DOWNLOADING TENDER FORM	19 th October 2020
LAST DATE & TIME FOR SUBMISSION OF TENDERS	20 th October 2020, till 12:00 Hrs.
DATE OF OPENING OF TECHNICAL BIDS	20 th October 2020, 15:00 Hrs.
<u>TENTATIVE DATE</u> OF OPENING OF PRICE BIDS	WILL BE INTIMATED LATER TO THE TECHNICALLY QUALIFIED/ RESPONSIVE TENDERERS

All other contents of the tender document including terms & conditions remain unaltered.

Note: Prospective Bidders are also advised to check the website regularly prior to the Closing date and time for submission of bids.

Sd/-
Stores Officer
For Secretary-cum-Scientific Director